

SUPERVISOR TEST

ORANGETOWN TOWN
POLICE DEPT.
DATAMASTER dmt: 115206

Date: 04/26/2025

OPERATOR NAME:
T AMBROSE
Permit #: 5343525040
Operator Agency/Dept: ORANGETOWN TOWN P.D.
Reference Standard #: 25040

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:48
INTERNAL STANDARD	VERIFIED	21:48
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	21:49
BLANK TEST	0.00	21:50
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	21:50
BLANK TEST	0.00	21:51
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	21:51
BLANK TEST	0.00	21:52

Making a false statement which a person
does not believe to be true is a Class A
misdemeanor in violation of Section 210.45
of the penal law of the State of New York.

Date 4/26/25
Signature [Signature]

