

SUPERVISOR TEST

ORANGETOWN TOWN  
POLICE DEPT.  
DATAMASTER dmt: 115206

Date: 08/15/2025

OPERATOR NAME:  
T AMBROSE  
Permit #: 53435  
Operator Agency/Dept: ORANGETOWN TOWN P.D.  
Reference Standard #: 25170

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:09
INTERNAL STANDARD	VERIFIED	08:09
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	08:09
BLANK TEST	0.00	08:10
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	08:10
BLANK TEST	0.00	08:11
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	08:12
BLANK TEST	0.00	08:12

Making a false statement which a person does not believe to be true is a Class A misdemeanor in violation of Section 210.45 of the penal law of the State of New York.

Date 8/15/25  
Signature T Ambrose

