

SUPERVISOR TEST

ORANGETOWN TOWN
POLICE DEPT.
DATAMASTER dmt: 115206

Date: 12/22/2025

OPERATOR NAME:
T AMBROSE
Permit #: 53435
Operator Agency/Dept: ORANGETOWN TOWN P.D.
Reference Standard #: 25250

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:48
INTERNAL STANDARD	VERIFIED	21:48
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.06	21:49
BLANK TEST	0.00	21:49
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.06	21:50
BLANK TEST	0.00	21:51
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.06	21:51
BLANK TEST	0.00	21:52

Making a false statement which a person does not believe to be true is a Class A misdemeanor in violation of Section 210.45 of the penal law of the State of New York.

Date 12/22/25
Signature T. Ambrose

