

SUPERVISOR TEST

ORANGETOWN TOWN
POLICE DEPT.
DATAMASTER dmt: 115206

Date: 02/12/2026

OPERATOR NAME:
T AMBROSE
Permit #: 53435
Operator Agency/Dept: ORANGETOWN TOWN P.D.
Reference Standard #: 25390

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:28
INTERNAL STANDARD	VERIFIED	11:28
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	11:28
BLANK TEST	0.00	11:29
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	11:29
BLANK TEST	0.00	11:30
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	11:31
BLANK TEST	0.00	11:31

Making a false statement which a person does not believe to be true is a Class A misdemeanor in violation of Section 210.45 of the penal law of the State of New York.

Date 2/12/26
Signature T. Ambrose

