

SUPERVISOR TEST

ORANGETOWN TOWN
POLICE DEPT.
DATAMASTER dmt: 115206

Date: 07/10/2025

OPERATOR NAME:
T AMBROSE
Permit #: 53435
Operator Agency/Dept: ORANGETOWN TOWN P.D.
Reference Standard #: 25110

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:12
INTERNAL STANDARD	VERIFIED	16:12
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	16:12
BLANK TEST	0.00	16:13
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	16:14
BLANK TEST	0.00	16:14
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	16:15
BLANK TEST	0.00	16:16

Making a false statement which a person does not believe to be true is a Class A misdemeanor in violation of Section 210.45 of the penal law of the State of New York.

Date 7/10/25

Signature [Signature] 258

