

SUPERVISOR TEST

ORANGETOWN TOWN
POLICE DEPT.
DATAMASTER dmt: 115206

Date: 05/03/2025

OPERATOR NAME:
T AMBROSE
Permit #: 53435
Operator Agency/Dept: ORANGETOWN TOWN P.D.
Reference Standard #: 25040

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:30
INTERNAL STANDARD	VERIFIED	12:30
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	12:30
BLANK TEST	0.00	12:31
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	12:31
BLANK TEST	0.00	12:32
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	12:32
BLANK TEST	0.00	12:33

Making a false statement which a person does not believe to be true is a Class A misdemeanor in violation of Section 210.45 of the penal law of the State of New York.

Date 5/3/25

Signature *T. Ambrose*

