

SUPERVISOR TEST

ORANGETOWN TOWN
POLICE DEPT.
DATAMASTER dmt: 115206

Date: 05/10/2025

OPERATOR NAME:
T AMBROSE
Permit #: 53435
Operator Agency/Dept: ORANGETOWN TOWN P.D.
Reference Standard #: 25040

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:08
INTERNAL STANDARD	VERIFIED	22:08
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	22:08
BLANK TEST	0.00	22:09
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	22:10
BLANK TEST	0.00	22:11
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	22:11
BLANK TEST	0.00	22:12

Making a false statement which a person does not believe to be true is a Class A misdemeanor in violation of Section 210.45 of the penal law of the State of New York.

Date 5/10/25

Signature T. Ambrose

