

SUPERVISOR TEST

ORANGETOWN TOWN
POLICE DEPT.
DATAMASTER dmt: 115206

Date: 10/21/2025

OPERATOR NAME:
T AMBROSE
Permit #: 53435
Operator Agency/Dept: ORANGETOWN TOWN P.D.
Reference Standard #: 25110

Nominal = 0.10 g/210L

| | | |
|-------------------------|----------|-------|
| BLANK TEST | 0.00 | 16:48 |
| INTERNAL STANDARD | VERIFIED | 16:49 |
| REFERENCE STANDARD TEMP | 34.02c | |
| REFERENCE STANDARD | 0.04 | 16:49 |
| BLANK TEST | 0.00 | 16:50 |
| REFERENCE STANDARD TEMP | 34.02c | |
| REFERENCE STANDARD | 0.04 | 16:50 |
| BLANK TEST | 0.00 | 16:51 |
| REFERENCE STANDARD TEMP | 34.02c | |
| REFERENCE STANDARD | 0.04 | 16:51 |
| BLANK TEST | 0.00 | 16:52 |

Making a false statement which a person
does not believe to be true is a Class A
misdemeanor in violation of Section 210.45
of the penal law of the State of New York.

Date: 10/21/25
Signature: [Handwritten Signature]

